



DIVISION OF INDUSTRY SERVICES
Plumbing Product Review
P.O. Box 7302
Madison, Wisconsin 53701-2658
TTY: Contact Through Relay

Scott Walker, Governor
Dave Ross, Secretary

October 24, 2016

DENTSPLY SIRONA
KARL NITTINGER
221 W PHILADELPHIA STREET, STE 60W
YORK PA 17401

SIRONA DENTAL INCORPORATED
JIM LEKOVICH
4835 SIRONA DR SUITE 100
CHARLOTTE NC 28273

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: SIRONA DENTAL INCORPORATED
Product Name: (trans id 2782323) SIRONA DENTAL TREATMENT CENTERS
Model Number(s): INTEGO D3543 AND TENEO D3509
Product File No: 20160255

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of October 2021.

This approval is contingent upon compliance with the following stipulation(s):

- All potable water lines connected to these devices shall incorporate cross connection control suitable for high hazard.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter
Environmental Engineer - Plumbing Product Reviewer
Department of Safety and Professional Services
Division of Industry Services
Bureau of Technical Services
(608) 267-1401 Phone
(608) 267-9723 Fax
glen.schlueter@wi.gov E-mail